

LASER INSPECTOR L1000/D4000L

ポータブルバーコードTRAD + ISO D/bility検査機

THE BARCODE VERIFIER THAT PERFORMS
THE TRADITIONAL METHOD & THE PARTIAL ISO SCAN REFLECTANCE PROFILE METHOD



■ 特徴

- ハンドヘルドレーザスキャナを採用
- 非接触読取で曲面や凸凹面上のバーコード検査に最適
- デコーダビリティ値(ISO/ANSI規格)及びJISx0501等Traditional規格に照合した検査
- 3信号音による表示[合格/注意/警告/不合格]
- マルチスキャンで総合検査判定機能
- コンテナスモードで読取率検査
- バー幅偏差をグラフと5つのLEDで表示

製品仕様

供給電源	単3アルカリ乾電池×4 *充電兼用AC/DCコンバータ	読取コード	EAN/JAN/UPC CODE 39
光源波長	半導体レーザー 赤色光(670nm)		ANSI 3OF9
分解能	0,20mm		ANSI 3OF9/WITH MOD43
スキャン回数	60スキャン/秒		AIAG
表示	LCD 16文字×2行		HIBC
寸法(L×W×H)	198,1×116,8×48,3(mm)		LOGMARS
重量	約454g		Interleaved 2 of 5
専用プリンタ	*TP40Ammonite サーマルプリンタ		Case Code
	* オプション		CODABAR
			EAN/UCC Code128(D4000Lのみ)
			USS Code128
			Code 93(L1000のみ)

L1000 検査データ例

Laser Inspector
L1000 Rev. E
Single Scan Analysis
90-115% EAN13
Acceptable
49 02720 00775 7
Modck: 7 Pass
Expect: 7

D/bility .48 C
-100% Tol. +100%

-----RRARRRR+++++

D4000L 検査データ例

Inspector D4000L
Revision A.02
Cont. Scan Analysis
UCC/EAN-128
Warning PCS
AI(00)

F01357913574
C102468024652
Decodes 011/011
%Decodes 100%
Modck:5 042 Pass
Expect:5 042

D/bility .55 B

-100% Tol. +100%

-----AR+++++++



The bar code industry's
most intelligent choice
since 1975

MUNAZO CO., LTD

The reliable partner of RJS & PRINTRONIX
RJS & PRINTRONIX
Department of Product Management
URL <http://munzao.jp>
<http://munazo.us>(English)
E-Mail munazo@cam.hi-ho.ne.jp

〒658-0032

神戸市東灘区向洋町中 6-9
KOBE FM
phone: 078-857-5447
fax: 078-857-5443

"As AutoScan II manufacturer since 2001"